

QUALITY ASSURANCE AUDIT REPORT

North Texas Commission Ambient Air and Meteorological Monitoring

Prepared for:

**North Texas Commission
8445 Freeport Parkway
Irving, TX 75063**

Prepared by:

**AECOM
13640 Briarwick Drive (78729)
P.O. Box 201088
Austin, TX 78720-1088**

**Conducted:
November 2024**

[AECOM](#)
[P.O. Box 201088](#)
[Austin, TX 78720-1088](#)
[13640 Briarwick Drive](#)
[Austin, TX 78729](#)

EXECUTIVE SUMMARY

On November 11th through 13th and November 18th through 20th, an audit team from the AECOM ambient air group in Austin, Texas conducted performance and technical system audits of the North Texas Commission (NTC) ambient air monitoring network. The audits provide an independent assessment of the monitoring program.

The monitoring program at NTC consists of continuous gas chromatographs (GC), volatile organic compound (VOC) canister collection systems, and meteorological sensors including wind speed, wind direction, and temperature.

The performance and technical systems audit results indicate acceptable responses for measurement systems with the exceptions summarized below.

The aspirator fan at Kennedale was non-operational at the time of the audit. The fan was replaced on November 13th.

The bearings on the wind speed sensor at Dish were outside of audit specifications (<0.3 g/cm) for starting threshold (0.3 g-cm CW, 0.4 g-cm CCW). The bearings were replaced on November 16th.

The bearings on the wind speed sensor at Benbrook were outside of audit specifications (<0.3 g/cm) for starting threshold (0.3 g-cm CW, 0.4 g-cm CCW). The bearings were replaced on November 14th.

The wind direction sensor at Dish was outside of audit specifications (<2.0° error) for alignment (5.6° error) with a maximum total error of 7.1°. Wind direction sensor was re-aligned to 0.5° error with a maximum total error of 2.0°.

The wind direction sensor at Rushing was outside of audit specification (<2.0° error) for alignment (2.3° error) with a maximum total error of -2.6°. Wind direction sensor was re-aligned to -11.8° error with a maximum total error of -2.4°.

The wind direction sensor at Mineral Wells was outside of audit specification (<2.0° error) for alignment (27.7° error) with a maximum total error of 28.8°. Wind direction sensor was re-aligned to 0.0° error with a maximum total error of 1.1°.

The shelter at Wichita Falls has a leak at the top of the shelter which may cause dirt and water damage if not quickly addressed.

The logbook entries at Wichita Falls does not portray the VOC schedule accurately. A cannister listed on the logbook was not found at the shelter. The operator addressed this concern by adding a clarifying logbook entry.

There are trees and tall vegetation within 20 meters of the shelter at Mineral Wells, Decatur, Lancaster, Weatherford, Keller and Rhome. This can negatively impact the accuracy of the data collected by the sensors and gas chromatograph due to air flow obstruction and skewing local measurements.

The sample cane at Weatherford and Gainesville is dirty and requires cleaning to ensure accurate data.

Logbook entries for all site requires departure time.

Out of the 48 compounds being analyzed, six compounds (Ethylene, Acetylene, Styrene, n-Decane, 1,2,3-Trimethylbenzene, and n-Undecane) were found to be outside of the audit objective of 70% - 130% recovery at several sites. In addition, the Elm Fork, Decatur and Rhome sites had the following GC compound recoveries outside of the audit specification:

Locations	Compounds
Elm Fork	Ethane
Decatur	2-Methylhexane
Rhome	M&P-Xylene

These network GC audit results are comparable historically to other AECOM auto-GC audits. The GC audit results are contained in table ES-1. Technical systems audit results demonstrate satisfactory operational procedures for collecting valid data.

A performance evaluation (PE) sample is prepared by the AECOM QA group on a quarterly basis and submitted to the VOC laboratory for analysis. This performance evaluation sample contained known (spiked) concentrations of the target VOCs. A review of the sample recoveries for the spiked target VOCs shows that only four out of the forty-four compounds were not within the range of expected values (70-130%).

- 1,2,4-Trimethylbenzene (51.3%)
- 1,3,5-Trimethylbenzene (65.8%)
- 1-Hexene (Not Detected)
- 4-Ethyltoluene (65.1%)

AECOM QA staff shared the performance evaluation results with the VOC laboratory, and no other corrective action was taken. We will continue to evaluate the compounds in our PE samples and work with the lab to resolve this discrepancy. GD Air's most recent performance evaluation canister results for the third quarter of 2024 are contained below in Table ES-2.

Table ES-1. Audit Standard Results for all Network GCs

Compound Name	CAS Number	Audit Conc (ppbc)	Benbrook		Decatur		Dish		Eagle Mountain Lake	
			Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery
Ethane	74-84-0	51.80	39.9	77.0%	42.3	81.6%	47.4	91.5%	48.6	93.7%
Ethylene	74-85-1	17.29	7.0	40.3%	13.7	79.0%	15.0	86.9%	16.1	93.4%
Propane	74-98-6	12.85	11.2	87.2%	11.1	86.3%	12.2	94.9%	11.9	92.9%
Propylene	115-07-1	12.97	9.9	76.0%	10.8	83.5%	11.9	91.9%	11.4	87.9%
Iso-Butane	75-28-5	17.13	16.7	97.6%	17.7	103.3%	18.5	107.8%	17.6	103.0%
N-Butane	106-97-8	17.62	17.9	101.5%	18.4	104.5%	19.5	110.6%	19.0	107.8%
Acetylene	74-86-2	8.65	4.7	54.5%	5.0	57.9%	5.6	65.3%	7.3	84.5%
Trans-2-Butene	624-64-6	17.29	16.8	96.9%	18.1	104.6%	19.1	110.7%	18.3	105.8%
1-Butene	106-98-9	17.62	16.7	94.6%	18.0	102.3%	18.8	106.9%	18.0	102.2%
Cis-2-Butene	590-18-1	17.29	17.6	101.9%	18.9	109.0%	20.0	115.7%	19.1	110.6%
Cyclopentane	287-92-3	21.82	21.1	96.8%	22.2	101.7%	23.6	108.1%	23.0	105.4%
Iso-Pentane	78-78-4	22.22	21.8	98.1%	22.9	103.3%	24.2	108.8%	23.6	106.4%
N-Pentane	109-66-0	22.22	21.8	98.1%	23.1	103.9%	24.3	109.4%	23.8	106.9%
1,3-Butadiene	106-99-0	16.97	16.5	97.0%	17.2	101.4%	17.8	104.7%	17.8	105.0%
Trans-2-Pentene	646-04-8	22.02	21.3	96.6%	22.3	101.1%	22.0	99.9%	23.5	106.5%
1-Pentene	109-67-1	22.22	19.5	87.8%	20.0	90.2%	18.6	83.5%	20.9	94.1%
Cis-2-Pentene	627-20-3	21.01	17.6	84.0%	18.4	87.5%	16.2	76.9%	19.6	93.1%
2,2-Dimethylbutane	75-83-2	26.18	24.8	94.7%	25.6	97.7%	23.6	90.2%	27.1	103.7%
2-Methylpentane	107-83-5	26.42	24.2	91.6%	25.7	97.3%	29.0	109.8%	26.4	100.1%
Isoprene	78-79-5	21.62	20.8	96.4%	15.9	73.7%	17.1	78.9%	17.4	80.7%
n-Hexane	110-54-3	25.94	24.2	93.3%	20.5	79.1%	23.8	91.9%	30.1	115.9%
Methylcyclopentane	108-87-2	26.18	20.8	79.6%	18.5	70.5%	20.4	78.1%	21.1	80.6%
2,4-Dimethylpentane	108-08-7	30.55	28.2	92.4%	27.7	90.8%	29.3	96.0%	31.8	104.0%
Benzene	71-43-2	25.70	20.6	80.2%	20.2	78.8%	21.6	83.9%	24.0	93.2%
Cyclohexane	110-82-7	25.94	25.2	97.0%	22.5	86.7%	23.5	90.7%	26.5	102.2%
2-Methylhexane	591-76-4	29.41	22.0	74.8%	20.5	69.6%	22.1	75.2%	24.0	81.5%
2,3-Dimethylpentane	565-59-3	29.70	30.5	102.8%	27.9	93.9%	29.3	98.6%	33.4	112.3%
3-Methylhexane	589-34-4	29.98	27.3	90.9%	25.2	84.2%	25.8	85.9%	30.4	101.3%
2,2,4-Trimethylpentane	540-84-1	34.26	30.3	88.6%	27.4	79.9%	30.8	89.8%	33.6	98.0%
n-Heptane	142-82-5	29.70	24.7	83.3%	23.7	79.7%	25.4	85.6%	28.9	97.2%
Methylcyclohexane	108-87-2	29.41	26.1	88.8%	23.9	81.3%	26.6	90.3%	28.1	95.4%
2,3,4-Trimethylpentane	565-75-3	33.94	29.1	85.8%	26.7	78.8%	30.9	90.9%	32.1	94.5%
Toluene	108-88-3	29.70	23.7	79.7%	24.4	82.2%	25.3	85.3%	28.9	97.5%
2-Methylheptane	592-27-8	33.94	27.01	79.6%	26.3	77.6%	29.4	86.5%	31.8	93.6%
3-Methylheptane	589-81-1	34.26	27.82	81.2%	26.6	77.5%	29.9	87.4%	32.3	94.2%
n-Octane	111-65-9	33.62	28.42	84.6%	26.9	79.9%	29.7	88.3%	32.1	95.6%
Ethylbenzene	100-41-4	33.29	25.45	76.5%	25.3	75.9%	27.9	83.9%	29.2	87.7%
M&P-Xylene	108-38-3	66.59	48.32	72.6%	51.4	77.2%	54.1	81.3%	56.7	85.1%
Styrene	100-42-5	32.00	19.72	61.6%	19.7	61.4%	22.9	71.5%	24.2	75.7%
O-Xylene	95-47-6	33.62	27.46	81.7%	26.0	77.5%	29.2	86.8%	30.8	91.7%
N-Nonane	111-84-2	36.73	30.18	82.2%	29.3	79.7%	32.8	89.4%	35.0	95.2%
Isopropylbenzene	98-82-8	36.73	27.44	74.7%	27.2	74.0%	31.4	85.4%	32.9	89.6%
n-Propylbenzene	103-65-1	35.27	25.65	72.7%	25.3	71.8%	29.0	82.3%	31.3	88.8%
1,3,5-Trimethylbenzene	108-67-8	36.73	26.52	72.2%	27.9	75.9%	29.6	80.5%	33.2	90.5%
1,2,4-Trimethylbenzene	95-63-6	36.00	27.56	76.6%	25.7	71.4%	29.6	82.1%	31.8	88.3%
n-Decane	124-18-5	40.40	27.50	68.1%	27.5	68.1%	31.9	78.9%	34.1	84.3%
1,2,3-Trimethylbenzene	526-73-8	34.91	22.83	65.4%	22.1	63.3%	26.2	75.0%	27.6	79.1%
n-Undecane	1120-21-4	41.33	25.76	62.3%	24.0	58.1%	27.2	65.7%	27.3	65.9%

^a Compound order based on elution time.

Table ES-1. (Continued) Audit Standard Results for all Network GCs

Compound Name	CAS Number	Audit Conc (ppbc)	Elm Fork		Everman		Flower Mound		Godley	
			Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery
Ethane	74-84-0	51.80	34.8	67.2%	51.0	98.5%	39.8	76.9%	50.2	96.9%
Ethylene	74-85-1	17.29	9.4	54.6%	17.8	103.1%	13.4	77.4%	16.2	93.9%
Propane	74-98-6	12.85	13.1	101.7%	12.7	98.7%	11.2	87.5%	12.4	96.4%
Propylene	115-07-1	12.97	11.4	88.2%	12.4	95.9%	10.0	77.1%	10.5	80.8%
Iso-Butane	75-28-5	17.13	19.1	111.2%	18.5	107.9%	15.9	92.7%	17.5	102.1%
N-Butane	106-97-8	17.62	19.8	112.7%	19.4	109.9%	16.7	95.0%	18.3	104.1%
Acetylene	74-86-2	8.65	8.7	100.9%	8.4	96.7%	7.3	84.9%	8.5	98.7%
Trans-2-Butene	624-64-6	17.29	19.5	112.7%	18.4	106.6%	16.4	94.6%	17.6	101.5%
1-Butene	106-98-9	17.62	19.0	108.1%	18.4	104.6%	15.9	90.4%	17.6	100.0%
Cis-2-Butene	590-18-1	17.29	20.1	116.4%	19.2	111.2%	16.9	97.9%	18.4	106.2%
Cyclopentane	287-92-3	21.82	24.0	110.1%	23.0	105.3%	20.5	93.9%	20.1	92.0%
Iso-Pentane	78-78-4	22.22	24.7	111.2%	23.7	106.6%	20.8	93.8%	23.2	104.6%
N-Pentane	109-66-0	22.22	24.4	109.7%	23.5	105.6%	21.3	95.9%	23.2	104.6%
1,3-Butadiene	106-99-0	16.97	18.7	110.1%	18.0	106.2%	15.2	89.4%	17.4	102.7%
Trans-2-Pentene	646-04-8	22.02	24.1	109.4%	23.2	105.6%	20.2	91.7%	22.4	101.8%
1-Pentene	109-67-1	22.22	23.9	107.4%	22.6	101.8%	18.6	83.8%	22.4	100.9%
Cis-2-Pentene	627-20-3	21.01	20.0	95.1%	20.0	95.1%	15.5	73.9%	19.8	94.4%
2,2-Dimethylbutane	75-83-2	26.18	28.4	108.5%	26.4	100.9%	23.5	89.9%	26.6	101.6%
2-Methylpentane	107-83-5	26.42	28.0	106.0%	25.8	97.5%	23.5	88.9%	26.2	99.1%
Isoprene	78-79-5	21.62	18.6	86.3%	18.8	87.2%	24.4	113.0%	18.1	83.6%
n-Hexane	110-54-3	25.94	26.2	100.9%	25.6	98.6%	24.4	94.2%	24.6	94.7%
Methylcyclopentane	108-87-2	26.18	26.9	102.6%	25.3	96.6%	19.9	76.0%	22.8	87.0%
2,4-Dimethylpentane	108-08-7	30.55	30.5	99.9%	32.5	106.4%	29.2	95.4%	32.9	107.8%
Benzene	71-43-2	25.70	23.9	92.8%	23.7	92.4%	20.1	78.0%	24.6	95.8%
Cyclohexane	110-82-7	25.94	26.5	102.3%	26.4	101.8%	22.5	86.7%	27.1	104.6%
2-Methylhexane	591-76-4	29.41	28.6	97.4%	26.3	89.4%	22.0	74.8%	26.0	88.4%
2,3-Dimethylpentane	565-59-3	29.70	31.0	104.3%	31.9	107.5%	28.5	95.9%	33.6	113.1%
3-Methylhexane	589-34-4	29.98	30.1	100.5%	29.8	99.4%	25.7	85.8%	31.0	103.5%
2,2,4-Trimethylpentane	540-84-1	34.26	35.0	102.3%	34.8	101.7%	30.3	88.5%	34.5	100.7%
n-Heptane	142-82-5	29.70	29.8	100.2%	29.6	99.7%	25.8	86.8%	29.7	99.9%
Methylcyclohexane	108-87-2	29.41	30.5	103.5%	30.3	103.0%	26.0	88.6%	29.5	100.3%
2,3,4-Trimethylpentane	565-75-3	33.94	33.7	99.2%	34.6	101.9%	29.5	86.9%	34.3	101.0%
Toluene	108-88-3	29.70	27.6	92.9%	28.9	97.5%	25.2	84.9%	28.9	97.2%
2-Methylheptane	592-27-8	33.94	33.2	97.7%	33.3	98.1%	28.4	83.6%	33.0	97.3%
3-Methylheptane	589-81-1	34.26	32.8	95.8%	33.9	98.8%	28.8	84.1%	33.7	98.3%
n-Octane	111-65-9	33.62	33.6	100.0%	34.1	101.4%	28.6	85.1%	34.6	103.0%
Ethylbenzene	100-41-4	33.29	31.3	93.9%	31.3	93.9%	26.4	79.1%	29.5	88.6%
M&P-Xylene	108-38-3	66.59	61.1	91.8%	61.3	92.1%	50.2	75.3%	57.6	86.5%
Styrene	100-42-5	32.00	27.5	85.8%	26.1	81.5%	22.0	68.6%	26.6	83.2%
O-Xylene	95-47-6	33.62	32.2	95.6%	33.0	98.2%	26.9	80.0%	31.8	94.5%
N-Nonane	111-84-2	36.73	36.1	98.2%	37.6	102.5%	30.9	84.1%	36.5	99.4%
Isopropylbenzene	98-82-8	36.73	34.4	93.8%	36.0	98.0%	28.6	77.7%	32.2	87.8%
n-Propylbenzene	103-65-1	35.27	32.6	92.5%	33.9	96.2%	26.8	75.9%	31.4	89.0%
1,3,5-Trimethylbenzene	108-67-8	36.73	33.3	90.5%	36.6	99.7%	25.7	70.1%	32.2	87.6%
1,2,4-Trimethylbenzene	95-63-6	36.00	34.2	95.1%	34.8	96.8%	26.6	73.9%	35.3	98.0%
n-Decane	124-18-5	40.40	37.7	93.3%	37.2	92.1%	29.3	72.4%	34.1	84.4%
1,2,3-Trimethylbenzene	526-73-8	34.91	32.5	93.1%	31.5	90.3%	22.5	64.5%	27.6	79.1%
n-Undecane	1120-21-4	41.33	37.4	90.5%	30.3	73.2%	25.4	61.5%	33.5	80.9%

^a Compound order based on elution time.

Table ES-1. (Continued) Audit Standard Results for all Network GCs

Compound Name	CAS Number	Audit Conc (ppbc)	Kennedale		Mansfield		Rhome		Rushing		UTA	
			Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery
Ethane	74-84-0	51.80	50.4	97.3%	41.8	80.7%	51.5	99.4%	43.9	84.8%	54.2	104.6%
Ethylene	74-85-1	17.29	17.2	99.2%	13.5	78.2%	16.9	97.6%	13.3	77.2%	14.3	82.6%
Propane	74-98-6	12.85	12.5	97.5%	11.6	90.1%	12.2	94.6%	11.5	89.8%	13.8	107.1%
Propylene	115-07-1	12.97	12.1	93.4%	9.8	75.8%	11.0	85.0%	9.7	74.7%	10.7	82.8%
Iso-Butane	75-28-5	17.13	16.1	94.2%	18.0	105.3%	16.8	97.9%	17.6	102.7%	19.8	115.5%
N-Butane	106-97-8	17.62	16.3	92.8%	18.9	107.1%	17.5	99.6%	18.5	104.8%	20.9	118.6%
Acetylene	74-86-2	8.65	8.3	96.6%	6.5	74.6%	7.7	89.3%	6.1	71.0%	7.9	91.8%
Trans-2-Butene	624-64-6	17.29	15.8	91.4%	17.9	103.5%	17.1	99.1%	17.8	102.9%	20.2	117.1%
1-Butene	106-98-9	17.62	16.7	94.8%	17.6	99.8%	16.9	96.0%	17.6	99.9%	19.9	112.8%
Cis-2-Butene	590-18-1	17.29	19.8	114.8%	18.6	107.7%	17.7	102.1%	18.6	107.8%	20.9	120.9%
Cyclopentane	287-92-3	21.82	19.8	91.0%	22.7	104.1%	21.3	97.8%	22.4	102.7%	25.9	118.5%
Iso-Pentane	78-78-4	22.22	21.0	94.8%	23.3	104.7%	22.4	100.6%	23.5	106.0%	26.3	118.5%
N-Pentane	109-66-0	22.22	21.1	95.3%	23.5	105.5%	22.3	100.5%	23.4	105.4%	26.0	116.9%
1,3-Butadiene	106-99-0	16.97	15.8	93.1%	17.5	103.0%	16.8	98.9%	17.7	104.2%	19.8	116.7%
Trans-2-Pentene	646-04-8	22.02	20.6	93.5%	22.5	102.0%	21.6	98.0%	23.0	104.3%	25.5	115.8%
1-Pentene	109-67-1	22.22	20.6	92.8%	22.4	101.0%	21.4	96.1%	23.2	104.2%	25.5	114.8%
Cis-2-Pentene	627-20-3	21.01	18.2	86.6%	18.8	89.3%	18.5	87.9%	20.2	96.0%	22.0	104.8%
2,2-Dimethylbutane	75-83-2	26.18	24.5	93.5%	27.1	103.7%	25.1	95.7%	27.1	103.7%	30.6	116.7%
2-Methylpentane	107-83-5	26.42	23.9	90.6%	26.2	99.2%	24.2	91.7%	25.4	96.1%	29.2	110.6%
Isoprene	78-79-5	21.62	17.3	79.9%	16.4	76.0%	16.8	77.7%	16.8	77.6%	18.6	86.2%
n-Hexane	110-54-3	25.94	24.9	96.0%	24.3	93.6%	27.6	106.4%	22.4	86.2%	28.9	111.5%
Methylcyclopentane	108-87-2	26.18	22.9	87.4%	23.1	88.2%	21.9	83.7%	20.2	77.2%	26.7	102.2%
2,4-Dimethylpentane	108-08-7	30.55	32.8	107.4%	28.5	93.4%	32.2	105.5%	30.4	99.6%	34.6	113.3%
Benzene	71-43-2	25.70	24.9	97.1%	23.6	91.8%	21.7	84.3%	22.9	89.2%	27.6	107.3%
Cyclohexane	110-82-7	25.94	26.4	101.9%	24.8	95.5%	24.2	93.2%	24.8	95.5%	30.2	116.6%
2-Methylhexane	591-76-4	29.41	27.0	91.8%	25.4	86.4%	25.1	85.2%	23.1	78.6%	28.6	97.1%
2,3-Dimethylpentane	565-59-3	29.70	32.8	110.6%	29.2	98.2%	31.3	105.3%	30.5	102.8%	36.3	122.3%
3-Methylhexane	589-34-4	29.98	31.9	106.4%	28.0	93.3%	28.7	95.8%	27.8	92.7%	33.5	111.7%
2,2,4-Trimethylpentane	540-84-1	34.26	34.8	101.5%	31.1	90.7%	34.0	99.2%	30.6	89.3%	37.5	109.5%
n-Heptane	142-82-5	29.70	30.3	102.0%	27.0	90.9%	29.6	99.6%	26.5	89.1%	32.5	109.4%
Methylcyclohexane	108-87-2	29.41	29.7	101.2%	26.8	91.1%	28.8	97.8%	26.4	89.6%	32.6	110.8%
2,3,4-Trimethylpentane	565-75-3	33.94	34.3	101.0%	30.5	89.8%	33.8	99.6%	29.8	87.7%	36.4	107.3%
Toluene	108-88-3	29.70	29.6	99.7%	26.6	89.6%	28.6	96.4%	27.6	92.8%	31.0	104.5%
2-Methylheptane	592-27-8	33.94	34.1	100.5%	29.5	87.0%	33.4	98.3%	29.8	87.9%	35.2	103.8%
3-Methylheptane	589-81-1	34.26	34.6	101.1%	30.1	88.0%	34.0	99.1%	30.2	88.2%	36.1	105.5%
n-Octane	111-65-9	33.62	35.3	105.2%	29.9	89.1%	34.9	103.8%	29.7	88.4%	36.0	107.1%
Ethylbenzene	100-41-4	33.29	32.0	96.0%	26.7	80.3%	31.1	93.3%	27.8	83.6%	31.6	95.0%
M&P-Xylene	108-38-3	66.59	63.1	94.8%	51.0	76.7%	38.8	58.3%	54.4	81.8%	61.7	92.7%
Styrene	100-42-5	32.00	28.5	89.2%	24.0	74.9%	26.3	82.3%	25.1	78.4%	28.5	89.0%
O-Xylene	95-47-6	33.62	33.1	98.6%	26.7	79.5%	32.6	96.9%	29.6	88.1%	33.1	98.4%
N-Nonane	111-84-2	36.73	38.2	104.0%	31.8	86.5%	38.1	103.8%	33.7	91.8%	39.4	107.4%
Isopropylbenzene	98-82-8	36.73	34.8	94.7%	27.8	75.6%	35.1	95.7%	30.5	83.0%	34.7	94.4%
n-Propylbenzene	103-65-1	35.27	34.1	96.8%	27.3	77.5%	33.3	94.4%	29.3	83.2%	33.3	94.3%
1,3,5-Trimethylbenzene	108-67-8	36.73	34.7	94.6%	28.0	76.1%	33.2	90.4%	31.1	84.6%	33.4	90.9%
1,2,4-Trimethylbenzene	95-63-6	36.00	36.2	100.6%	30.5	84.8%	35.1	97.4%	33.0	91.8%	36.4	101.2%
n-Decane	124-18-5	40.40	39.2	97.0%	33.2	82.2%	37.9	93.9%	32.7	80.9%	36.8	91.1%
1,2,3-Trimethylbenzene	526-73-8	34.91	31.2	89.3%	26.7	76.5%	29.4	84.2%	26.7	76.6%	30.4	87.2%
n-Undecane	1120-21-4	41.33	37.8	91.5%	36.7	88.9%	29.5	71.5%	32.1	77.6%	40.1	97.1%

^a Compound order based on elution time.

Table ES-2. Results of Performance Standard for Off-Site Analytical Lab

Compound Name	CAS Number	Input Concentration	Lab Results	Percent Recovery
1,1,1-Trichloroethane	71-55-6	3.5	3.4	98.7%
1,1,1,2-Tetrachloroethane	79-34-5	3.5	2.9	82.1%
1,1,2-Trichloroethane	79-00-5	3.5	3.4	97.4%
1,1-Dichloroethane	75-34-3	3.4	3.5	102.2%
1,1-Dichloroethene	75-35-4	3.5	3.5	98.6%
1,2,4-Trimethylbenzene	95-63-6	3.4	1.8	51.3%
1,2-Dibromoethane	106-93-4	3.5	3.3	93.1%
1,2-Dichloroethane	107-06-2	3.4	3.3	95.5%
1,2-Dichloropropane	78-87-5	3.5	3.2	91.7%
1,3,5-Trimethylbenzene	108-67-8	3.5	2.3	65.8%
1,3-Butadiene	106-99-0	7.0	7.7	109.9%
1-Butene	106-98-9	3.5	3.3	93.1%
1-Hexene	592-41-6	3.2	ND	ND
1-Pentene	109-67-1	3.5	3.4	95.7%
2,2,4-Trimethylpentane	540-84-1	3.5	3.7	106.3%
4-Ethyltoluene (p-Ethyltoluene)	622-96-8	3.5	2.3	65.1%
Benzene	71-43-2	3.5	3.3	94.6%
Bromomethane	74-83-9	3.4	3.3	98.2%
c-1,3-Dichloropropene	10061-01-5	3.4	3.1	92.4%
Carbon tetrachloride	56-23-5	3.5	3.4	98.3%
Chlorobenzene	108-90-7	3.5	3.0	85.8%
Chloroform	67-66-3	3.4	3.3	97.0%
Chloromethane (Methyl Chloride)	74-87-3	3.5	3.2	91.2%
Cyclohexane	110-82-7	3.5	3.1	87.4%
Dichlorodifluoromethane (Freon-12)	75-71-8	3.5	3.3	93.8%
Ethane	74-84-0	20.9	20.1	96.0%
Ethene	74-85-1	6.8	7.2	105.4%
Ethylbenzene	100-41-4	3.6	2.7	75.1%
Ethylene Chloride (Dichloromethane)	75-09-2	3.5	3.4	98.7%
m-Xylene & p-Xylene	106-42-3+108-38-3	7.0	6.0	85.3%
n-Butane	106-97-8	3.5	3.3	95.8%
n-Heptane	142-82-5	3.5	3.4	96.6%
n-Hexane	110-54-3	10.5	10.5	99.9%
n-Pentane	109-66-0	3.5	3.4	96.6%
o-Xylene	95-47-6	3.5	2.9	82.4%
Propane	74-98-6	3.3	3.5	105.9%
Propylene	115-07-1	7.0	6.6	93.3%
Styrene	100-42-5	3.5	2.5	70.5%
t-1,3-Dichloropropene	10061-02-6	3.4	3.6	106.9%
Tetrachloroethene	127-18-4	3.4	2.9	85.6%
Toluene	108-88-3	3.5	2.9	82.4%
Trichloroethene	79-01-6	3.4	3.4	100.2%
Trichlorofluoromethane (Freon-11)	75-69-4	3.5	3.1	88.6%
Vinyl Chloride	75-01-4	3.4	3.3	94.7%